



Vascutherm

Please call one of the following numbers **FIRST** for any problems with you VascuTherm Machine. From 8 am to 5 pm call (877) 242-3232 or call (214)-502-8800 which is answered 24 hours a day. If they are unable to correct the problem, please contact your representative (REP). These numbers are for **TECHNICAL SUPPORT ONLY**. There are separate numbers if you need to contact your (REP), you will find their numbers listed at the bottom of this page.

Step #1

Place machine above the patient's body level.

Plug the power cord in the back and turn the power switch on (It's just above the power cord).

Plug the 2 large ends of the water hose into the back of the machine (the pointed ends). It does not matter which hole you plug the large ends into.

Plug the small one into the small hole. (The ends that plug in the back of the machine can not plug into the wrap and visa versa)

Step #2

Plug the wrap in (it doesn't matter which large end you plug into which. These are flat ends of the hose.) Plug in the end that you have leftover.

Step #3

If you are given "DVT CALF SLEEVES."

Plug them into the back of the machine just to the right of where you plugged the cold wrap into (there is no left or right with these sleeves so it doesn't matter which one, left or right, that you plug into; or what leg that they you put them on).

Push them in and turn slightly to the right, you will hear or feel them click in place.

Step #4

PAY ATTENTION TO YOUR MACHINE!

If the green screen does not read "VASCUTHERM READY" OR "OFF" follow these steps:

1. Hold the **COOL** button down until you hear a beep.
2. Press the **COMP** button. If it reads "COMPRESSION ALTERNATING" then hold the **COMP** button down until you hear a beep and the screen reads "COMPRESSION OFF."
3. If it reads "COMPRESSION OFF" then you don't have to worry about doing anything else.
4. This step just resets the timer for 1 hour.

All machines
are a 10 day
rental

All machines
are to be
returned to
Dr. Powell's
office **NOT** to
the surgery
center
or hospital



Step #5 (Turning the cool on)

Place the designated wrap on the patient (the dimpled side against the patient).

Hold the COLD button down until you hear a "BEEP" (The cold will run for 1 hour. If you would like to use the COLD for more than 1 hour just hold the COLD button down again).

Step #6 (Working with your dvt sleeves)

Press then hold the CALF button. The screen will say "DVT CALF BILATERAL".

(This part will not work if you did not follow STEP #4).

All machines use distilled water only!

Low coolant flow means add water under clear cap.

Throw away all cold wraps and dvt sleeves after your last use, including the small clear hoses connected to the dvt sleeves!

If you have problems with your vascuthem unit call the numbers listed at the top of this information sheet (technical support). If you have called technical support and they were unable to take care of your problem, call your rep-- chris kelley @ (479)-409-2889 or keith farlow @ (479)-841-9299. All calls will be returned within 24 hours.

If you have any problems or questions, please feel free to call our office at (479) 582-4647 between the hours of 8:30 a.m. and 4:30 p.m., Monday – Friday. If you have an emergency outside of our normal office hours, go to the nearest emergency room or call (479) 433-6116.